sasan
medical disposable products

Kod/Code: SVCXXXXXZ

www.sasan.com.tr

VAKUM (VAVD) KONTROL CİHAZI

VAVD Kontrol Cihazı , vakum destekli venöz drenajının hassas ve güvenli olarak sağlanması için özel olarak tasarlanmıştır. Venöz drenajındaki kan akımı yeterli olmadığında rezervuarda negatif basınç oluşturarak yeterli kan akışını sağlamak için vakum destekli venöz drenaj kullanılır.

ÖZELLİKLERİ

- **Analog manometre göstergeli**
  Vakum kaynağı ve rezervuardaki basıncı gösteren 2 ayrı manometresi bulunmaktadır.

- **Hafif ve ergonomik tasarım**

- **Hassas ve güvenli vakum kontrolü**
  Kardiyotomi rezervuarına uygulanan vakum güvenilir, izlenebilir ve hassastır.

- **VAVD standartlarına uygun**
  Vakum destekli venöz drenaj, hastadan venöz dönüşü artırmak için venöz rezervuarı kapağına özel bir vakum kontrol cihazının bağlanması ile elde edilir.

- **Açma-kapama anahtarı**
  Basınç tahliyesi sayesinde güvenilirdir.

- **Venöz rezervuarı vakum ayarlama düğmesi**
  Venöz rezervuara uygulanan vakumun sürekli olarak doğru olmasını sağlar ve uygulanan vakum seviyesini kolayca düzenler.

- **Tutucu**
  Askılığa monte etmek için tutucu mevcuttur .

- **1/4 " giriş ve çıkış bağlantıları**
  Cihaz ile rezervuar ve vakum kaynağı arasında doğru ve kolay bağlantı sağlar.

KATALOG NO	AÇIKLAMA
SVC820002	Vakum (VAVD) Kontrol Cihazı

Özellikler	SVC820002 Vakum (VAVD) Kontrol Cihazı
Ölçüm Aralığı	-760 mmHg ile 0 mmHg ve -120 mmHg ile 0 mmHg, ± % 3 hassasiyet
Vakum Girişi	¼" (6 mm)
Vakum Çıkışı	¼" (6 mm)
Ana Gövde	Alüminyum Gövde
Açma Kapama	Hassas Ayarlı Açma ve Kapama
Boyut (Y x G x D)	231 mm x 184 mm x 108 mm
Ağırlık	2,4 kg

Katalog No: SVCXXXXXZ
S: Sasan Kodlama Başlangıcı
VC: Vakum Kontrol Cihazı
XXXXX:Kodu
Z: 2-İki manometreli, 1-Tek manometreli

CE
1984

SKTL-01/05/03.08.2021

Sasan Sağlık Malzemeleri Üretim ve Pazarlama A.Ş. / Dağyaka Mah. 2004.Cad. No:6 KahramanKazan 06980 Ankara,Turkey / Tel: +90 312 815 46 50

sasan
medical disposable products

Kod/Code: SVCXXXXXZ

www.sasan.com.tr

VACUUM (VAVD) CONTROL DEVICE

Vacuum (VAVD) Control Device has been specifically designed to ensure that vacuum assised venous drainage is precise and safe. When the blood flow in venous drainage is not sufficient, vacuum-assisted venous drainage is used to provide sufficient blood flow by creating negative pressure in the reservoir.

FEATURES

- **With an analog manometer gauge**
There are 2 manometers showing the pressures at vacuum source and the reservoir sides.

- **Light and ergonomic design**

- **Precise and safe vacuum control**
The suction applied to the cardiotomy reservoir is reliable , traceable and consistent.

- **Comply with related VAVD standards**
Vacuum assisted venous drainage is achieved by connecting a special vacuum control device to venous cover of reservoir increase venous return from the patient.

- **On-off switch**
It works reliably with pressure relief.

- **Vacuum regulating button of venous reservoir**
Ensures that the vacuum applied to the venous reservoir is consistently correct and easily regulates applied vacuum level.

- **Holder**
Vacuum Control Device has a holder for mounting on the IV pole.

- **1/4 " inlet and outlet connections**
It provides an accurate and easy connection between the device and the reservoir and vacuum source.

Specifications	SVC820002 Vacuum (VAVD) Control Device
Measuring Range	-760 mmHg to 0 mmHg and -120 mmHg to 0 mmHg, ± 3% accuracy
Vacuum Inlet	¼" (6 mm)
Vacuum Outlet	¼" (6 mm)
Body	Aluminum Body
On/Off	Fine tuning On/Off
Size (H x W x D)	231mmx184mmx108mm
Weight	2,4 kg

MATERIAL CODE	DESCRIPTION
SVC820002	Vacuum (VAVD) Control Device

Material Code: SVCXXXXXZ
S: Sasan Begining of Code
VC: Vacuum Control Device
XXXXX: Code Number
Z: 2-With two manometers, 1-Single manometer

CE
1984

SKTL-01/05/03.08.2021

Sasan Sağlık Malzemeleri Üretim ve Pazarlama A.Ş. / Dağyaka Mah. 2004.Cad. No:6 KahramanKazan 06980 Ankara,Turkey / Tel: +90 312 815 46 50